SCHLÖDER
EMV-Systeme & Komponenten

CDN calibration set

To calibrate a CDN the following items are required:

- **2 x CDG A 31xx** (**specific calibration** adaptor for CDN)
- **2 x CDG A 3100** (**fastening angle** and **150/50 Ohm adaptor**)

Two fastening angle and 150/50 Ohm adaptor should be ordered for the first CDN. For each following CDN only two specific adaptors has to be ordered.



| < --- CDG A 31xx ----> | < ---------- CDG A 3100 ---------- > |



50 Ohm
termination

311705 DB_CDN_e

Distributed by: Reliant EMC LLC, 3311 Lewis Ave, Signal Hill CA 90755, 408-916-5750, www.reliantemc.com



Insertion loss 150 / 50 Ohm adaptor (two in series)

Alu mounting plate (thick 2.5 mm)



311705 DB_CDN_e